



THE HASHEMITE KINGDOM OF
JORDAN
Accreditation Unit



Annex (1)
Updated on: 30-10-2018

To the Accreditation Certificate No. JAS Cal. – 004 Dated 2018-03-22
For Laboratory at Rum for Calibration

Scope of Accreditation

Calibration of Temperature

Measurand	Measuring Range	Calibration and measurement Capability (CMC)/ Uncertainty of measurement	Calibration Methods/ Standards/ Remarks
Temperature Calibrators	-80 °C ≤ T < -50 °C	0.59 °C	Comparison with Standard RTD SOP-T012, rev.1/a EURAMET Calibration Guide No. 13 Version 4.0 (09/2017) Revision date:03/01/2018 Effective date:07/01/2018
	-50 °C ≤ T < -25 °C	0.26 °C	
	-25 °C ≤ T < 0 °C	0.26 °C	
	0 °C ≤ T < 25 °C	0.26 °C	
	25 °C ≤ T < 50 °C	0.3 °C	
	50 °C ≤ T < 100 °C	0.3 °C	
	100 °C ≤ T < 200 °C	0.38 °C	
	200 °C ≤ T < 300 °C	0.57 °C	
300 °C ≤ T < 400 °C	0.66 °C		
Temperature Controllers	-80 °C ≤ T < -50 °C	0.59 °C	Comparison with Standard RTD SOP-T012, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018
	-50 °C ≤ T < -25 °C	0.26 °C	
	-25 °C ≤ T < 140 °C	0.5 °C	Using Dry Block SOP-T039, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018
	150 °C ≤ T < 300 °C	0.41 °C	Using Dry Block SOP-T039, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018
	300 °C ≤ T < 400 °C	0.58 °C	



THE HASHEMITE KINGDOM OF
JORDAN
Accreditation Unit



Annex (1)
Updated on: 30-10-2018

To the Accreditation Certificate No. **JAS Cal. – 004** Dated **2018-03-22**
For Laboratory at Rum for Calibration

Scope of Accreditation

Calibration of Temperature

Measurand	Measuring Range	Calibration and measurement Capability (CMC)/ Uncertainty of measurement	Calibration Methods/ Standards/ Remarks
Thermometers/ RTD thermometer	-80 °C ≤ T < -50 °C -50 °C ≤ T < -25 °C	0.59 °C 0.26 °C	Comparison with Standard RTD SOP-T055, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018
	-25°C ≤ T < 140 °C	0.5 °C	Using Dry Block SOP-T054, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018
	150°C ≤ T < 300 °C 300°C ≤ T < 400 °C	0.41 °C 0.58 °C	Using Dry Block & Standard RTD SOP-T054 , rev.1/a Revision date:03/01/2018 Effective date:07/01/2018
Freezer / Refrigerators	-25°C ≤ T < 8 °C	1.0 °C	Using Thermometer SOP-069, rev.1/a EURAMET cg-8Version 2.1 (10/2011) Revision date:03/01/2018 Effective date:07/01/2018



THE HASHEMITE KINGDOM OF
JORDAN
Accreditation Unit



Annex (1)
Updated on: 30-10-2018

To the Accreditation Certificate No. **JAS Cal. – 004** Dated **2018-03-22**
For Laboratory at Rum for Calibration

Scope of Accreditation

Calibration of Temperature

Measurand	Measuring Range	Calibration and measurement Capability (CMC)/ Uncertainty of measurement	Calibration Methods/ Standards/ Remarks
Thermo Couples / Temperature Sensors with Indicators	-80 °C ≤ T < -50 °C -50 °C ≤ T < -25 °C	0.58 °C 0.26 °C	Comparison with Standard RTD SOP-T055, rev.1/a EURAMET cg-8Version 2.1 (10/2011) Revision date:03/01/2018 Effective date:07/01/2018
	-25°C ≤ T < 140 °C	0.5 °C	Using Dry Block & Standard RTD SOP-T130, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018
	150°C ≤ T < 300 °C 300°C ≤ T < 400 °C	0.41 °C 0.58 °C	Using Dry Block & Standard RTD SOP-T130, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018
Totally immersed Liquid in Glass Thermometer	-25°C ≤ T < 140 °C	0.5 °C	Using Dry Block SOP-T054, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018
	150°C ≤ T < 400 °C	1.0 °C	Using Dry Block & Standard RTD SOP-T054, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018



THE HASHEMITE KINGDOM OF
JORDAN
Accreditation Unit



Annex (1)
Updated on: 30-10-2018

To the Accreditation Certificate No. **JAS Cal. – 004** Dated **2018-03-22**
For Laboratory at Rum for Calibration

Scope of Accreditation

Calibration of Temperature

Measurand	Measuring Range	Calibration and measurement Capability (CMC)/ Uncertainty of measurement	Calibration Methods/ Standards/ Remarks
Data Loggers with External Sensors	-80 °C ≤ T < -50 °C -50 °C ≤ T < -25 °C	0.59 °C 0.26 °C	Comparison with Standard RTD SOP-T055, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018
	-25 °C ≤ T < 140 °C	0.5 °C	Using Dry Block & Standard RTD SOP-T130, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018
	150 °C ≤ T < 300 °C 300 °C ≤ T < 400 °C	0.41 °C 0.58 °C	Using Dry Block & Standard RTD SOP-T130, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018
Incubators	25 °C ≤ T < 50 °C 50 °C ≤ T < 70 °C	1.0 °C 0.6 °C	Using Thermometer SOP-T136, rev.1/a EURAMET/cg-20/v.01 Revision date:03/01/2018 Effective date:07/01/2018
Autoclaves	108 °C ≤ T < 134 °C	0.50 °C	Using Dry Block SOP-T070, rev.1/a Revision date:03/01/2018 Effective date:07/01/2018



THE HASHEMITE KINGDOM OF
JORDAN
Accreditation Unit



Annex (1)
Updated on: 30-10-2018

To the Accreditation Certificate No. JAS Cal. – 004 Dated 2018-03-22
For Laboratory at Rum for Calibration

Scope of Accreditation

Calibration of Temperature

Measurand	Measuring Range	Calibration and measurement Capability (CMC)/ Uncertainty of measurement	Calibration Methods/ Standards/ Remarks
Temperature Gauges	$-10^{\circ}\text{C} \leq T < 140^{\circ}\text{C}$	0.50 °C	Using Dry Block SOPT012, rev.1/a EURAMET Calibration Guide No. 13 Version 4.0 (09/2017) Revision date:03/01/2018 Effective date:07/01/2018
Ovens	$50^{\circ}\text{C} \leq T < 100^{\circ}\text{C}$	0.6 °C	Using Thermometer SOP-T029, rev.1/a EURAMET/cg-20/v.01 Revision date:03/01/2018 Effective date:07/01/2018
	$100^{\circ}\text{C} \leq T < 300^{\circ}\text{C}$	1 °C	Using Thermometer SOP-T029, rev.1/a EURAMET/cg-20/v.01 Revision date:03/01/2018 Effective date:07/01/2018
Liquid baths	$-10^{\circ}\text{C} \leq T < -25^{\circ}\text{C}$ $-25^{\circ}\text{C} \leq T < 0^{\circ}\text{C}$ $0^{\circ}\text{C} \leq T < 25^{\circ}\text{C}$ $25^{\circ}\text{C} \leq T < 50^{\circ}\text{C}$ $50^{\circ}\text{C} \leq T < 100^{\circ}\text{C}$	1.0 °C 1.0 °C 1.0 °C 0.4 °C 0.6 °C	Using Thermometer SOP-T030, rev.1/a EURAMET/cg-20/v.01 Revision date:03/01/2018 Effective date:07/01/2018
	$100^{\circ}\text{C} \leq T < 120^{\circ}\text{C}$	0.38 °C	Using Standard RTD Thermometer SOP-T030, rev.1/a EURAMET/cg-20/v.01 Revision date:03/01/2018 Effective date:07/01/2018



THE HASHEMITE KINGDOM OF
JORDAN
Accreditation Unit



Annex (1)
Updated on: 30-10-2018

To the Accreditation Certificate No. JAS Cal. – 004 Dated 2018-03-22
For Laboratory at Rum for Calibration

Scope of Accreditation

Calibration of Temperature

List of employees in the laboratory who are technically responsible for issuing the calibration certificates in the scope of accreditation:

1. Husam Hamdan: Technical Manger
2. Murshed Al-deiri: Lab Supervisor
3. Huda Shaltaf: Quality Manager
4. Bashar Drouf: Calibration and Qualification Engineer
5. Fares Younis : Calibration and Qualification Engineer