



THE HASHEMITE KINGDOM OF
JORDAN

Accreditation Unit



Annex (1)

**To The Accreditation Certificate No. JAS Test - 091 Dated 04-09-2018
for the Laboratory at I.D. Technical Services L.L.C.**

**Scope of Accreditation
In the Field of Environmental Testing**

Tested Parameter/ Type of Test/ Measured Quantity	Test Methods/ Standards
Sound level in work environment (continious measurment)	<ul style="list-style-type: none">Integrating sound level meter/ ISO 1996-2: 2017
Illuminance level in work environment (continious measurment)	<ul style="list-style-type: none">Silicon photodiode sensor / CNS 5119:1988-10
Flue gas (O ₂ ,CO,NO _x ,SO ₂ , CO ₂) (continious measurment)	<ul style="list-style-type: none">Electrochemical cells or infrared/ JS 1189:2006
Particulate matter (PM10.PM2.5, TSP) In work environment (continious measurment)	<ul style="list-style-type: none">Light scattering method/ OSHA technical manual section II chapter 3 (TED 01-00-015A)

List of employees in the laboratory who are technically responsible for issuing the test reports in the scope of accreditation:

- 1- General Manager: Eng. Hasan Alkhatib**
- 2- Deputy General Manager: Eng. Jawad Ramadan**
- 3- Quality Manager: Eng. Yazan Alqadi**
- 4- Technical Manager: Eng. Murad Sako**

الملحق رقم (١)

لشهادة الاعتماد رقم **JAS Test – 091** الممنوحة بتاريخ ٢٠١٨/٠٩/٠٤

لمختبر شركة الأيدي للخدمات الفنية ذ.م.م.

مجال الاعتماد

فحوصات البيئة

المواصفات المتبعة/ طرق الفحص	القيمة المقاسة / نوع الفحص/ الخاصية المقاسة
جهاز قياس شدة الصوت / ISO 1996-2: 2017	مستوى الضجيج في بيئة العمل (قياس مستمر)
جهاز قياس شدة الإضاءة / CNS 5119:1988-10	شدة الإضاءة في بيئة العمل (قياس مستمر)
الخلايا الكهروكيميائية أو الأشعة تحت الحمراء / JS 1189:2006	انبعاثات المداخل للغازات (الأكسجين وأول أكسيد الكربون وأكاسيد النيتروجين وثاني أكسيد الكبريت وثاني أكسيد الكربون) (قياس مستمر)
تشتت الضوء / OSHA technical manual section II chapter 3 (TED 01-00-015A)	الجسيمات العالقة (الجسيمات الكلية العالقة TSP والجسيمات الدقيقة (PM10,PM2.5) (قياس مستمر)

قائمة بالأشخاص الذين يتحملون المسؤولية الفنية لتقارير الاختبار الصادرة عن المختبر في مجال الاعتماد:

١. المدير العام / م. حسن الخطيب
٢. نائب المدير العام / م. جواد رمضان
٣. مدير الجودة / م. يزن القاضي
٤. المدير الفني / م. مراد ساكو