



Annex (1)

Updated on: 01-10-2019

To the Accreditation Certificate No. **JAS Cal. – 005** Dated **2018-05-08**

For the Laboratory of **SAMA for Calibration and Qualification**

Scope of Accreditation

Calibration of Temperature and Pressure (Onsite / Permanent)

Measurand	Measuring Range	Calibration and Measurement Capability (CMC) ^a	Calibration Methods/ Standards/ Remarks
Temperature			
Thermocouples	-40 °C to +140 °C	0.25 °C	Kaye low temperature reference (LTR)/Comparison with reference RTD Calibration is done according to EURAMET cg-08 SP14-1, Issue 2, revision 5, Issue date 10/08/2017, effective date: 30/07/2019
	-23 °C to +150 °C	0.5 °C	Jofra advanced temperature calibrator(ATC 156)/Comparison with reference RTD Calibration is done according to EURAMET cg-08 SP14-1, Issue 2, revision 5, Issue date 10/08/2017, effective date: 30/07/2019
	-30 °C to +150 °C	0.95 °C	Jofra reference/professional temperature calibrator (RTC 156)/Comparison with reference RTD Calibration is done according to EURAMET cg-08 SP14-1, Issue 2, revision 5, Issue date 10/08/2017, effective date: 30/07/2019
	50 to < 150 °C	0.62 °C	Kaye high temperature reference (HTR)/Comparison with reference RTD Calibration is done according to EURAMET cg-08 SP14-1, Issue 2, revision 5, Issue date 10/08/2017, effective date: 30/07/2019
	150 to < 300 °C	0.62 °C	
	300 to 400 °C	0.62 °C	



Annex (1)

Updated on: 01-10-2019

To the Accreditation Certificate No. **JAS Cal. – 005** Dated **2018-05-08**

For the Laboratory of **SAMA for Calibration and Qualification**

Scope of Accreditation

Calibration of Temperature and Pressure (Onsite / Permanent)

Measurand	Measuring Range	Calibration and measurement capability (CMC) ^a	Calibration Methods/ Standards/ Remarks
Temperature			
Thermocouples	50 to < 150 °C	1.1 °C	Jofra advanced temperature calibrator(ATC 650) Comparison with reference RTD Calibration is done according to EURAMET cg-08 SP14-1, Issue 2, revision 5, Issue date 10/08/2017, effective date: 30/07/2019
	150 to < 250 °C	1.1 °C	
	250 to < 350 °C	1.1 °C	
	350 to 650 °C	5.1 °C	
Total immersion Glass Thermometers	-20 °C to +110 °C	0.15 °C	Fluke Micro Bath Calibration is done according to SP14-3, Issue 2, revision 6, Issue date 10/08/2017, effective date: 28/09/2019
Data Loggers	5 °C to < 25 °C	0.65 °C	Reference generator for calibration of Humidity and temperature instruments (Hydrogen 1A) Calibration is done according to EURAMET cg-08 SP14-4, Issue 2, revision 3, Issue date 10/08/2017, effective date: 12/06/2019
	25 °C to < 40 °C	0.65 °C	
	40 °C to +50 °C	0.65 °C	
	0 °C to < 25 °C	0.84 °C	Reference generator for calibration of Humidity and temperature instruments (Hydrogen 2A) Calibration is done according to EURAMET cg-08 SP14-4, Issue 2, revision 3, Issue date 10/08/2017, effective date: 12/06/2019
	25 °C to < 45 °C	0.84 °C	
	45 °C to +60 °C	0.84 °C	
Pressure			
Pressure	1 to < 100 Pa	1.2 Pa	Reference Differential Pressure Calibrator Druck DPI 610 Calibration is done according to EURAMET cg-17 SP14-5, Issue 2, revision 6, Issue date 10/08/2017, effective date: 30/07/2019
	100 to 15000 Pa	2.1 Pa	



Annex (1)

Updated on: 01-10-2019

To the Accreditation Certificate No. **JAS Cal. – 005** Dated **2018-05-08**

For the Laboratory of **SAMA for Calibration and Qualification**

Scope of Accreditation

Calibration of Temperature and Pressure (Onsite / Permanent)

Measurand	Measuring Range	Calibration and measurement Capability (CMC) ^a	Calibration Methods/ Standards/ Remarks
Pressure			
Pressure	0.1 to 50 Bar	0.046 Bar	Pressure Calibrator Druck DPI 620 Pneumatic/Comparison with reference Pressure Sensor PM620 (100 Bar) Calibration is done according to EURAMET cg-17 SP14-5, Issue 2, revision 6, Issue date 10/08/2017, effective date: 30/07/2019
Pressure	0.1 to 50 Bar	0.28 Bar	Hydraulic Pump with Fluke Digital Pressure Gauge Calibration is done according to EURAMET cg-17 SP14-5, Issue 2, revision 6, Issue date 10/08/2017, effective date: 30/07/2019
	0.1 to 50 Bar	0.28 Bar	Hydraulic Pump with Martel Digital Pressure Gauge Calibration is done according to EURAMET cg-17 SP14-5, Issue 2, revision 6, Issue date 10/08/2017, effective date: 30/07/2019
Vacuum Pressure	15000 Pa to > -100 Pa	2.1 Pa	Reference Differential Pressure Calibrator Druck DPI 610 Calibration is done according to ISO/TS 3567 SP14-6, , Issue 2, revision 6, Issue date 10/08/2017, effective date: 30/07/2019
	-100 to to1 Pa	2.8 Pa	Hydraulic Pump with Fluke Digital Pressure Gauge Calibration is done according to ISO/TS 3567 SP14-6, , Issue 2, revision 6, Issue date 10/08/2017, effective date: 30/07/2019
	-0.8 to 0.1 Bar	0.28 Bar	Hydraulic Pump with Martel Digital Pressure Gauge Calibration is done according to ISO/TS 3567 SP14-6, , Issue 2, revision 6, Issue date 10/08/2017, effective date: 30/07/2019
	-0.8 to 0.1 Bar	0.016 Bar	Hydraulic Pump with Fluke Digital Pressure Gauge Calibration is done according to ISO/TS 3567 SP14-6, , Issue 2, revision 6, Issue date 10/08/2017, effective date: 30/07/2019

a) The reported CMCs are expressed at approximately the 95 % level of confidence, using a coverage factor of $k = 2$. The actual measurement uncertainty of a specific calibration performed by the laboratory may be greater than the CMC due to the behavior of the customer's device and to influences from the circumstances of the specific calibration.



Annex (1)

Updated on: 01-10-2019

To the Accreditation Certificate No. JAS Cal. – 005 Dated 2018-05-08

For the Laboratory of SAMA for Calibration and Qualification

Scope of Accreditation

Calibration of Temperature and Pressure (Onsite / Permanent)

List of employees in the laboratory who are technically responsible for issuing the calibration certificates in the scope of accreditation:

1. Malek Shamlawe / Technical Manager
2. Afnan Dawoud / Quality Manager
3. Essa Jamal/ Head of Lab
4. Anas Dahshan / Calibration and Qualification Supervisor
5. Ismael Al-Khateeb / Calibration and Qualification Supervisor
6. Mohammad Adnan / Calibration and Qualification Engineer

Voluntary Withdrawal - Full Scope Date: 01/01/2021